

Date of Report: September 19, 1990
 Date Submitted: September 14, 1990
 Project: 3600 90 Day Report + 3200 Metro Sump

RESULTS OF ANALYSES OF THE WATER SAMPLES
 FOR SELECTED METALS BY ICP
 Results Reported as mg/L (ppm)

Sample #	Cadmium (ppm)	Chromium (ppm)	Copper (ppm)	Nickel (ppm)	Lead (ppm)	Zinc (ppm)
3200 Monthly M16566A	*	0.46	0.14	0.38	*	0.05

Quality Assurance

Method Blank	<0.01	<0.01	<0.01	<0.01	<0.01	<0.01
M16566A (Duplicate)	<0.01	0.52	0.15	0.41	0.01	0.05
M16566A (Matrix Spike) Spiked @ 5 ppm						
Percent Recov.	105%	100%	81%	116%	98%	117%
M16566A (Matrix Spike Duplicate) Spiked @ 5 ppm						
Percent Recov.	90%	89%	69%	101%	80%	97%

* - Analysis for this analyte was not requested.